

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
ECBC												
Phase Ia												
SLS-B1	45.2	0.157	13-Aug-02	0.187	17.06%	1	1	0.8361	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	VC difference > 15%.
SLS-B2	40.4	0.140	27-Aug-02	0.385	3.88%	3	4	0.7841	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	Inadequate curve fit.
SLS-B3	38.6	0.134	27-Aug-02	0.410	0.04%	1	5	0.8376	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B4	33.3	0.116	28-Aug-02	0.288	15.91%	1	2	0.9378	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	VC difference > 15%.
SLS-B5	26.6	0.092	28-Aug-02	0.233	4.43%	2	4	0.8086	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	Inadequate curve fit.
SLS-B6 (25 ug/ml NR 1 hr)	39.5	0.137	4-Sep-02	0.255	7.59%	1	2	0.9621	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	R&D: 3 replicate ODs/concentration
SLS-B7 (50 ug/ml NR 1 hr)	39.1	0.136	4-Sep-02	0.330	3.18%	1	2	0.9749	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	R&D: 3 replicate ODs/concentration
SLS-B8 (25 ug/ml NR 3 hr)	36.5	0.126	4-Sep-02	0.508	3.64%	1	3	0.9639	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	R&D: 3 replicate ODs/concentration
SLS-B9 (50 ug/ml NR 3 hr)	33.1	0.115	4-Sep-02	0.457	1.39%	1	4	0.9678	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	R&D: 3 replicate ODs/concentration
SLS-B11	42.9	0.149	9-Sep-02	0.349	6.33%	1	2	0.9332	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B12	35.3	0.123	10-Sep-02	0.326	5.41%	1	3	0.9211	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B13	33.0	0.114	10-Sep-02	0.414	6.50%	1	4	0.8802	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B14 (33 ug/ml NR)	37.6	0.130	11-Sep-02	0.347	1.97%	1	3	0.9241	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B15 (33 ug/ml NR)	42.8	0.148	11-Sep-02	0.303	3.16%	1	1	0.8408	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	Inadequate curve fit.
SLS-B16 (33 ug/ml NR)	34.8	0.121	11-Sep-02	0.345	3.43%	1	2	0.9770	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B17 (33 ug/ml NR)	34.3	0.119	11-Sep-02	0.389	17.94%	0	4	0.8377	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	VC difference > 15%. No points between 10 & 50%.
SLS-B18	39.2	0.136	17-Sep-02	0.430	7.88%	1	2	0.9472	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B19	44.7	0.155	17-Sep-02	0.422	13.89%	1	1	0.9389	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B20	34.8	0.121	17-Sep-02	0.445	4.12%	1	3	0.9364	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B21	38.6	0.134	17-Sep-02	0.402	1.66%	1	3	0.8969	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B22	43.5	0.151	18-Sep-02	0.394	2.94%	1	1	0.9271	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
SLS-B23	39.7	0.138	18-Sep-02	0.423	1.71%	1	2	0.9253	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
SLS-B24	45.6	0.158	18-Sep-02	0.283	10.48%	0	2	0.8502	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	No points between 10 & 50%.
SLS-B25	44.6	0.155	18-Sep-02	0.311	13.03%	1	0	0.8784	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	No points between 50 & 90%.
Phase Ib												
ECBC-3T3-lb-01 SLS-P1	34.0	0.118	22-Jan-03	0.300	2.23%	1	3	0.9245	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
ECBC-3T3-lb-01 SLS-P2	31.3	0.109	22-Jan-03	0.214	2.18%	1	4	0.8744	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
ECBC-3T3-lb-02 SLS-P3	13.2	0.046	29-Jan-03	0.270	23.27%	2	3	0.8703	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	VC difference > 15%; IC50 out of range
ECBC-3T3-lb-03 SLS-P4	56.1	0.195	4-Feb-03	0.438	7.34%	1	2	0.8206	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	IC50 out of range
ECBC-3T3-lb-04 SLS-P5	43.0	0.149	25-Feb-03	0.750	3.31%	1	1	0.9827	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
ECBC-3T3-lb-05 SLS-P7	40.8	0.141	26-Feb-03	0.443	6.47%	1	1	0.9702	100, 68, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
ECBC-3T3-lb-06 SLS-P9	44.9	0.156	4-Mar-03	0.450	3.57%	1	1	0.9403	100, 68.0, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
ECBC-3T3-lb-07 SLS-P12	37.3	0.129	11-Mar-03	0.568	10.54%	1	4	0.9314	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
ECBC-3T3-lb-08 SLS-P13	47.2	0.164	18-Mar-03	0.517	6.58%	1	1	0.9566	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
Phase II												
SLS-P1	41.4	0.144	17-Jun-03	0.409	4.01%	3	3	0.9561	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P2	36.1	0.125	17-Jun-03	0.452	16.14%	3	4	0.9411	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	% VC difference > 15
SLS-P3	44.5	0.154	24-Jun-03	0.427	8.32%	3	3	0.9434	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P4	39.5	0.137	24-Jun-03	0.460	0.14%	3	4	0.9202	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P5	44.2	0.153	1-Jul-03	0.619	2.60%	3	4	0.9365	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P6	37.8	0.131	1-Jul-03	0.563	3.20%	2	4	0.9361	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P7	42.1	0.146	8-Jul-03	0.485	5.48%	1	5	0.9162	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P8	41.5	0.144	8-Jul-03	0.630	4.97%	2	4	0.9461	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P9	40.3	0.140	15-Jul-03	0.450	6.36%	1	5	0.9250	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
SLS-P10	35.2	0.122	15-Jul-03	0.629	4.12%	3	3	0.9751	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P11	38.7	0.134	22-Jul-03	0.488	3.70%	2	4	0.9769	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P12	39.1	0.136	22-Jul-03	0.554	1.92%	3	4	0.9760	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P13	41.6	0.144	29-Jul-03	0.700	0.18%	3	4	0.9440	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P14	40.7	0.141	29-Jul-03	0.730	3.11%	3	4	0.9663	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P15	43.2	0.150	5-Aug-03	0.649	0.59%	2	4	0.9591	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P16	44.1	0.153	6-Aug-03	0.276	3.23%	4	4	0.9790	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P17	37.3	0.129	31-Aug-03	0.710	5.38%	2	4	0.9482	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P18 sealer	32.4	0.112	31-Aug-03	0.545	4.39%	3	3	0.8897	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	R&D
SLS-P19	41.4	0.144	1-Sep-03	0.613	2.00%	3	3	0.9625	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P20	38.4	0.133	9-Sep-03	0.350	0.88%	3	4	0.9350	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P21	43.0	0.149	23-Sep-03	0.650	3.04%	2	4	0.9637	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P22	41.2	0.143	29-Oct-03	0.406	1.21%	3	4	0.9289	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P23	41.8	0.145	4-Nov-03	0.378	8.20%	4	4	0.9577	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P24	44.7	0.155	5-Nov-03	0.333	3.43%	4	3	0.9518	80, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	

Phase III

SLS-P1	37.5	0.130	13-Jan-04	0.355	3.82%	3	3	0.8860	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P2	34.9	0.121	13-Jan-04	0.442	8.96%	3	3	0.9641	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P3	40.8	0.142	21-Jan-04	0.461	4.62%	2	3	0.9751	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P4	29.4	0.102	21-Jan-04	0.511	3.62%	2	3	0.9672	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	IC50 out of range
SLS-P5	43.7	0.151	27-Jan-04	0.299	2.09%	3	4	0.9766	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P6	42.8	0.148	27-Jan-04	0.384	1.89%	2	3	0.9558	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P7	43.1	0.149	3-Feb-04	0.378	6.60%	4	4	0.9779	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
SLS-P8	38.5	0.134	3-Feb-04	0.379	7.38%	2	4	0.9662	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P9	38.5	0.134	10-Feb-04	0.375	8.36%	3	4	0.9315	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P10	35.9	0.124	10-Feb-04	0.374	3.25%	3	4	0.9640	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P11	40.5	0.140	24-Feb-04	0.297	2.83%	3	4	0.9554	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P12	37.3	0.129	24-Feb-04	0.334	0.02%	2	3	0.9665	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P13	39.3	0.136	25-Feb-04	0.385	0.30%	3	4	0.9624	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P14	37.9	0.132	25-Feb-04	0.422	5.43%	4	4	0.9561	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P15	44.7	0.155	2-Mar-04	0.526	3.85%	2	5	0.9840	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P16	41.9	0.145	2-Mar-04	0.605	0.29%	2	4	0.9739	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P17	38.9	0.135	3-Mar-04	0.453	7.56%	3	4	0.9496	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P18	35.5	0.123	3-Mar-04	0.522	0.59%	3	3	0.9404	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P19	41.3	0.143	9-Mar-04	0.539	7.29%	3	4	0.9586	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P20	37.7	0.131	9-Mar-04	0.535	0.73%	2	4	0.9731	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P21	42.7	0.148	16-Mar-04	0.563	0.59%	2	3	0.9849	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P22	38.9	0.135	16-Mar-04	0.548	0.03%	3	4	0.9759	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P23	43.4	0.150	23-Mar-04	0.632	3.43%	3	4	0.9714	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P24	42.1	0.146	23-Mar-04	0.707	2.19%	2	4	0.9858	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P25	52.7	0.183	30-Mar-04	0.667	2.75%	2	5	0.9661	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	IC50 out of range
SLS-P26	43.0	0.149	30-Mar-04	0.623	0.88%	3	3	0.9556	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P27	45.9	0.159	6-Apr-04	0.521	2.17%	2	4	0.9766	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P28	43.9	0.152	6-Apr-04	0.614	1.41%	3	4	0.9785	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P29	46.3	0.161	13-Apr-04	0.477	4.37%	3	5	0.9579	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
SLS-P30	43.1	0.149	13-Apr-04	0.609	1.67%	1	5	0.9420	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P31	44.1	0.153	20-Apr-04	0.473	5.99%	1	5	0.9456	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P32	39.4	0.136	20-Apr-04	0.481	2.79%	3	4	0.9762	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P33	44.8	0.155	27-Apr-04	0.434	8.49%	2	4	0.9548	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P34	42.1	0.146	27-Apr-04	0.448	8.96%	3	4	0.9624	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P35	49.3	0.171	4-May-04	0.611	1.23%	3	4	0.9828	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P36	42.4	0.147	4-May-04	0.680	4.09%	2	4	0.9626	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P37	44.8	0.155	11-May-04	0.588	2.31%	2	5	0.9713	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P38	43.2	0.150	11-May-04	0.682	3.69%	3	4	0.9645	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P39	37.8	0.131	18-May-04	0.418	7.64%	3	4	0.9578	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P40	37.0	0.128	18-May-04	0.408	1.70%	2	4	0.9541	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P41	45.0	0.156	25-May-04	0.506	2.77%	2	5	0.9772	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P42	42.1	0.146	25-May-04	0.575	1.65%	2	4	0.9733	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P43	42.8	0.148	15-Jun-04	0.698	6.20%	3	4	0.9689	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P44	42.2	0.146	15-Jun-04	0.695	8.92%	4	4	0.9648	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P45	45.9	0.159	22-Jun-04	0.561	1.81%	3	5	0.9718	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P46	46.1	0.160	22-Jun-04	0.650	1.33%	2	5	0.9772	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P47	40.2	0.139	29-Jun-04	0.421	8.18%	4	4	0.9603	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P48	37.6	0.130	29-Jun-04	0.468	10.36%	3	4	0.9512	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P49	40.2	0.139	13-Jul-04	0.325	12.65%	4	4	0.9524	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P50	NA	NA	20-Jul-04	0.414	4.06%	1	1	NA	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	IC50 out of range; IC50 > 51.6;
SLS-P51	NA	NA	20-Jul-04	0.414	16.20%	1	5	NA	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	IC50 out of range; IC50 > 51.6; % VC difference > 15;

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
SLS-P52	NA	NA	27-Jul-04	0.471	14.02%	3	1	NA	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	IC50 out of range; IC50 < 30.8;
SLS-P53	NA	NA	27-Jul-04	0.555	8.43%	5	1	NA	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	IC50 out of range; IC50 < 30.8;
SLS-P54	44.1	0.153	10-Aug-04	0.797	1.55%	3	5	0.9653	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P55	45.1	0.156	10-Aug-04	0.658	5.46%	3	4	0.9570	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P56	NA	NA	17-Aug-04	0.372	34.25%	2	5	NA	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	PC fails; % VC difference > 15
SLS-P57	40.4	0.140	17-Aug-04	0.523	6.59%	4	4	0.9579	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P58	47.1	0.163	24-Aug-04	0.477	4.19%	2	5	0.9215	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P59	40.6	0.141	24-Aug-04	0.462	7.30%	4	4	0.9589	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P60	53.7	0.186	31-Aug-04	0.754	3.56%	2	6	0.8457	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	IC50 out of range
SLS-P61	60.1	0.208	31-Aug-04	0.726	3.36%	2	6	0.9203	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	NO	IC50 out of range
SLS-P62	43.4	0.150	14-Sep-04	0.635	5.64%	2	5	0.9006	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P63	41.4	0.144	14-Sep-04	0.625	6.52%	2	5	0.9614	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P64	37.4	0.130	28-Sep-04	0.473	6.10%	3	4	0.9400	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P65	38.8	0.135	28-Sep-04	0.394	4.91%	3	4	0.9681	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P66	37.0	0.128	5-Oct-04	0.520	3.86%	2	4	0.9495	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P67	33.4	0.116	5-Oct-04	0.554	4.23%	3	3	0.9603	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P68	42.7	0.148	19-Oct-04	0.472	0.62%	2	5	0.9632	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P69	43.6	0.151	19-Oct-04	0.349	0.38%	1	5	0.9659	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P70	39.7	0.138	26-Oct-04	0.468	3.33%	3	4	0.9687	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P71	44.9	0.156	27-Oct-04	0.504	3.38%	2	3	0.9416	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P72	45.8	0.159	2-Nov-04	0.517	1.76%	3	5	0.9405	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	
SLS-P73	45.7	0.158	2-Nov-04	0.517	0.08%	2	5	0.9685	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
SLS-P74	46.6	0.161	16-Nov-04	0.510	0.42%	2	5	0.9461	80.0, 66.1, 54.6, 45.2, 37.3, 30.8, 25.5, 21.1	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
------------------------------	--------------------------------------	-----------	---	---	--	--	--	------------------------------------	-----------------------------------	---------------------------------	----------------------------------	----------------------------------

FAL**Phase Ia**

B1(1a/3T3/DF1/FA L/SLS)	53.9	0.187	3-Sep-02	0.402	11.18%	0	1	0.9577	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	No point between 10 & 50% viability
B2(1a/3T3/DF2/FA L/SLS)	NA	NA	3-Sep-02	0.419	15.17%	1	1	0.7691	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	Bad values for 6.3 ug/mL wells. VC difference > 15%.
B3(1a/3T3/DF3/FA L/SLS)	50.8	0.176	3-Sep-02	0.420	3.73%	0	1	0.9583	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	No point between 10 & 50% viability
B4(1a/3T3/DF4/FA L/SLS)	44.4	0.154	3-Sep-02	0.490	2.60%	1	1	0.9800	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	
B5(1a/3T3/DF5/FA L/SLS)	51.0	0.177	3-Sep-02	0.503	8.01%	0	1	0.9812	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	No point between 10 & 50% viability
B6(1a/3T3/DF6/FA L/SLS)	49.8	0.173	3-Sep-02	0.441	6.29%	1	0	0.9517	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	No point between 50 & 90% viability.
B7(1a/3T3/DF7/FA L/SLS)	54.2	0.188	4-Sep-02	0.408	5.64%	0	1	0.8134	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	No point between 10 & 50% viability
B8(1a/3T3/DF8/FA L/SLS)	50.2	0.174	4-Sep-02	0.337	34.90%	0	1	0.8010	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	VC difference > 15%. No point between 10 & 50% viability
B9(1a/3T3/DF9/FA L/SLS)	52.1	0.181	4-Sep-02	0.484	0.79%	0	1	0.9657	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	No point between 10 & 50% viability
B10(1a/3T3/DF10/ FAL/SLS)	52.5	0.182	4-Sep-02	0.459	7.20%	0	1	0.9389	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	No point between 10 & 50% viability
B11(1a/3T3/DF11/F AL/SLS)	46.4	0.161	4-Sep-02	0.509	6.94%	0	3	0.9422	400, 200, 100, 50, 25, 12.5, 6.3, 3.1	2.00	NO	No point between 10 & 50% viability
1a/3T3/DF14/FAL/ SLS	23.0	0.080	18-Sep-02	0.900	3.51%	1	3	0.8277	150, 102, 69.4, 47.2, 32.1, 21.9, 14.9, 10.1	1.47	NO	Inadequate curve fit.
1a/3T3/DF15/FAL/ SLS	46.7	0.162	18-Sep-02	0.547	7.61%	1	0	0.9736	150, 102, 69.4, 47.2, 32.1, 21.9, 14.9, 10.1	1.47	NO	No point between 50 & 90% viability
1a/3T3/DF16/FAL/ SLS	42.4	0.147	18-Sep-02	0.590	21.70%	1	0	0.9833	150, 102, 69.4, 47.2, 32.1, 21.9, 14.9, 10.1	1.47	NO	VC difference > 15%. No point between 50 & 90% viability.
1a/3T3/DF17/FAL/ SLS	46.6	0.161	18-Sep-02	0.442	4.00%	1	0	0.8646	150, 102, 69.4, 47.2, 32.1, 21.9, 14.9, 10.1	1.47	NO	No point between 50 & 90% viability
1a/3T3/DF18/FAL/ SLS	22.6	0.078	18-Sep-02	0.920	4.36%	2	3	0.8319	150, 102, 69.4, 47.2, 32.1, 21.9, 14.9, 10.1	1.47	NO	Inadequate curve fit.
1a/3T3/DF19/FAL/ SLS	23.1	0.080	18-Sep-02	0.936	4.30%	1	3	0.8350	150, 102, 69.4, 47.2, 32.1, 21.9, 14.9, 10.1	1.47	NO	Inadequate curve fit.
1a/3T3/DF28/FAL/ SLS	48.0	0.166	22-Oct-02	0.488	9.05%	0	1	0.9570	100, 68.0, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	No point between 10 & 50% viability
1a/3T3/DF29/FAL/ SLS	50.7	0.176	22-Oct-02	0.579	10.46%	0	3	0.8773	100, 68.0, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	No point between 10 & 50% viability

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
1a/3T3/DF30/FAL/SLS	42.0	0.146	23-Oct-02	0.768	6.31%	1	3	0.9433	100, 68.0, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	YES	
1a/3T3/DF31/FAL/SLS	46.8	0.162	23-Oct-02	0.795	2.60%	0	4	0.9321	100, 68.0, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	No point between 10 & 50% viability
1a/3T3/DF32/FAL/SLS	49.0	0.170	23-Oct-02	0.784	0.24%	0	1	0.9725	100, 68.0, 46.3, 31.5, 21.4, 14.6, 9.9, 6.7	1.47	NO	No point between 10 & 50% viability
1a3T3DF33FALSLS	48.9	0.169	30-Oct-02	0.676	2.03%	1	2	0.9532	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF34FALSLS	48.0	0.166	30-Oct-02	0.636	4.77%	1	2	0.9788	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF35FALSLS	48.7	0.169	30-Oct-02	0.684	2.23%	1	2	0.9811	100, 76.9, 59.2, 45.5, 35, 26.9, 20.7, 15.9	1.30	YES	
1a3T3DF36FALSLS	53.0	0.184	30-Oct-02	0.545	4.83%	1	1	0.8486	100, 76.9, 59.2, 45.5, 35, 26.9, 20.7, 15.9	1.30	NO	Inadequate curve fit.
1a3T3DF37FALSLS	50.8	0.176	31-Oct-02	0.660	1.09%	1	3	0.9261	100, 76.9, 59.2, 45.5, 35, 26.9, 20.7, 15.9	1.30	YES	
1a3T3DF38FALSLS ⁺	51.4	0.178	31-Oct-02	0.612	9.54%	1	4	0.9057	100, 76.9, 59.2, 45.5, 35, 26.9, 20.7, 15.9	1.30	YES	
1a3T3DF39FALSLS	51.3	0.178	31-Oct-02	0.630	0.19%	1	2	0.9749	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF40FALSLS	52.5	0.182	31-Oct-02	0.669	6.97%	1	1	0.9879	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF41FALSLS ⁺	47.1	0.163	5-Nov-02	0.581	3.57%	1	3	0.9757	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF42FALSLS	46.8	0.162	5-Nov-02	0.564	11.34%	1	3	0.9468	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF43FALSLS	36.6	0.127	6-Nov-02	0.649	6.40%	1	3	0.8929	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF44FALSLS ⁺	44.8	0.155	6-Nov-02	0.605	1.06%	2	3	0.9258	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF45FALSLS	40.7	0.141	12-Nov-02	0.618	0.88%	1	3	0.9756	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF46FALSLS	42.3	0.147	12-Nov-02	0.665	0.86%	1	3	0.9599	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF47FALSLS	42.1	0.146	12-Nov-02	0.674	3.71%	1	2	0.9811	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF48FALSLS	37.9	0.131	13-Nov-02	0.531	15.94%	2	3	0.8139	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	NO	VC difference > 15%.
1a3T3DF49FALSLS	38.7	0.134	13-Nov-02	0.561	14.96%	1	3	0.8648	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF50FALSLS	40.6	0.141	13-Nov-02	0.533	11.42%	2	3	0.9179	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF51FALSLS	40.3	0.140	20-Nov-02	0.689	0.29%	1	3	0.9478	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
1a3T3DF52FALSL S	42.5	0.147	20-Nov-02	0.780	1.37%	1	3	0.9682	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1a3T3DF53FALSL S	39.9	0.138	20-Nov-02	0.692	7.30%	2	3	0.9403	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
Phase Ib												
1b3T3CRT1FALSL S	34.4	0.119	4-Dec-02	0.618	16.76%	3	2	0.8479	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	NO	VC difference > 15%
1b3T3CTR2FALSL S	48.8	0.169	10-Dec-02	0.545	6.73%	1	2	0.9409	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1b3T3CTRFA LSL	24.5	0.085	17-Dec-02	0.453	1.97%	1	0	0.8653	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	NO	IC50 out of range; no points between 50 & 90% viability
1b3T3CTRFA LSL	43.5	0.151	7-Jan-03	0.597	2.23%	1	2	0.9631	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1b3T3CTRFA LSL	50.9	0.176	8-Jan-03	0.271	14.37%	1	1	0.9136	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	NR crystals in plate; stopped after 1 h
1b3T3CRTFA LSL	43.2	0.150	14-Jan-03	0.625	3.68%	1	3	0.9163	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1b3T3CRT2FALSL S	32.4	0.112	14-Jan-03	0.417	5.55%	1	2	0.9377	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1b3T3CRTFA LSL	70.1	0.243	15-Jan-03	0.432	2.31%	1	2	0.9000	82.6, 67.7, 56.0, 42.29, 38.25, 31.61, 26.13, 21.59	1.21	YES	IC50 out of range
1b3T3CRTFA LSL	35.3	0.122	21-Jan-03	0.651	1.86%	1	2	0.9727	100.00, 82.64, 68.30, 56.45, 46.65, 38.55, 31.86, 26.33	1.21	YES	
1b3T3CRTFA LSL	38.1	0.132	28-Jan-03	0.181	17.95%	1	0	0.9716	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	NO	NR crystals in plate; stopped after 1 h; VC difference > 15%; no point between 50 & 90% viability
1b3T3CRTFA LSL	58.7	0.204	29-Jan-03	0.646	8.07%	0	2	0.9573	100, 68.02, 46.28, 31.48, 21.42, 14.57, 9.91, 6.74		NO	No point between 10 & 50% viability
1b3T3CRTFA LSL	44.3	0.154	4-Feb-03	0.662	0.79%	1	1	0.9848	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
1b3T3CRTFA LSL	36.8	0.128	5-Feb-03	0.566	1.65%	1	1	0.9867	100, 82.645, 68.301, 56.447, 46.651, 38.554, 31.863, 26.333	1.21	YES	
1b3T3CRTFA LSL	48.0	0.166	26-Feb-03	0.310	15.17%	1	2	0.9457	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
Phase II												
A1SLS190603	49.1	0.170	17-Jun-03	1.031	2.49%	2	5	0.7802	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	NO	r2 too low

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
FAL.3T3.SLS2.A1.200603	54.6	0.189	18-Jun-03	0.684	6.26%	4	3	0.9851	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.A2.2 6.06.03	50.8	0.176	24-Jun-03	0.483	3.45%	3	4	0.9788	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.A2.2 7.06.03	50.7	0.176	25-Jun-03	0.564	0.19%	2	2	0.9878	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.B1.0 3.07.03	57.5	0.199	1-Jul-03	0.516	7.13%	1	4	0.9913	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	IC50 out of range
FAL.3T3.SLS.04.0 7.03	55.8	0.193	2-Jul-03	0.562	4.86%	4	3	0.9788	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	NO	IC50 out of range
FAL.3T3.SLS.10.0 7.03	52.5	0.182	8-Jul-03	0.640	0.86%	2	3	0.9794	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.11.07 .03	50.6	0.175	9-Jul-03	0.533	2.92%	2	3	0.9869	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.17.0 7.03	50.2	0.174	15-Jul-03	0.708	0.81%	2	3	0.9905	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.18.0 7.03	43.2	0.150	16-Jul-03	0.502	5.68%	2	3	0.9763	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.25.0 7.03	47.6	0.165	23-Jul-03	0.435	5.81%	1	2	0.9633	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.0708 03	30.5	0.106	5-Aug-03	0.725	0.11%	7	1	0.9204	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	NO	IC50 out of range
FAL.3T3.SLS.0808 03	36.2	0.126	6-Aug-03	0.463	1.17%	5	3	0.7811	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	NO	low r2
FAL.3T3.SLS.1209 03	39.4	0.137	10-Sep-03	0.768	4.53%	3	4	0.8322	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.1809 03	45.2	0.157	16-Sep-03	0.401	0.69%	4	3	0.9582	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.1909 03	45.0	0.156	17-Sep-03	0.377	0.62%	1	2	0.9790	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.2509 03	35.7	0.124	23-Sep-03	0.379	4.55%	3	2	0.9738	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.0310 03	51.2	0.178	1-Oct-03	0.596	5.23%	2	4	0.9344	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.1710 03	37.5	0.130	15-Oct-03	0.398	9.90%	3	2	0.9763	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.3010 03	49.8	0.173	28-Oct-03	0.310	12.63%	4	1	0.9702	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.3010 03 (should be 311003)	39.6	0.137	29-Oct-03	0.313	8.62%	3	3	0.9886	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
Phase III												
FAL.3T3.SLS.0801 04	55.0	0.191	6-Jan-04	0.615	0.20%	4	4	0.9771	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
FAL.3T3.SLS.09/0 1/04	53.3	0.185	7-Jan-04	0.592	7.04%	4	4	0.9727	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.15/0 1/04	67.0	0.232	13-Jan-04	0.841	1.98%	2	6	0.8901	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	NO	IC50 out of range
FAL.3T3.SLS.16/0 1/04	30.4	0.105	14-Jan-04	1.161	0.39%	6	2	0.8932	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.22/0 1/04	35.7	0.124	20-Jan-04	0.382	7.11%	3	2	0.9685	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL3T3.23-01-04	30.8	0.107	21-Jan-04	0.792	2.31%	2	2	0.9194	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL3T3.SLS.29-01- 04	41.4	0.144	27-Jan-04	0.467	0.43%	5	3	0.9671	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.29/0 1/04	44.3	0.153	28-Jan-04	0.453	1.44%	4	4	0.9721	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	YES	
FAL.3T3.SLS.5/02/ 04	26.9	0.093	3-Feb-04	0.417	2.14%	4	0	0.9317	100, 82.6, 68.5, 56.5, 46.7, 38.6, 31.9, 26.4	1.21	NO	recalculated values: IC50 out of range; no points between 50-100
FAL.3T3.SLS.06/0 2/04	38.8	0.135	4-Feb-04	0.427	4.23%	5	3	0.9136	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL3T3.SLS.25.02- .04	47.9	0.166	23-Feb-04	0.637	2.29%	3	4	0.9829	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.17/0 3/04	49.8	0.173	15-Mar-04	0.356	5.91%	4	3	0.9831	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.01/0 4/04	44.0	0.152	30-Mar-04	0.404	1.46%	2	2	0.9593	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.29/0 4/04	42.3	0.147	27-Apr-04	0.310	2.34%	3	5	0.9881	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.30/0 4/04	31.3	0.108	28-Apr-04	0.249	4.22%	6	1	0.9874	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.06/0 5/04	40.7	0.141	4-May-04	0.320	9.70%	2	3	0.9897	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.07/0 5/04	40.2	0.139	5-May-04	0.313	0.03%	3	3	0.9865	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.20/0 5/04	45.2	0.157	18-May-04	0.422	3.24%	2	3	0.9797	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.21.0 5.04	32.7	0.114	19-May-04	0.337	0.94%	2	2	0.9720	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.27/0 5/04	44.2	0.153	25-May-04	0.406	5.89%	3	3	0.9466	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.24.0 6.04	40.6	0.141	22-Jun-04	0.434	3.69%	4	3	0.9826	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.08.0 7.04	39.7	0.138	6-Jul-04	0.324	7.16%	2	3	0.9659	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.09.0 7.04	40.3	0.140	7-Jul-04	0.408	2.92%	2	3	0.9765	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
FAL.3T3.SLS.16.0 7.04	35.6	0.124	14-Jul-04	0.402	5.43%	2	2	0.9676	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.17.0 9.04	40.3	0.140	15-Sep-04	0.411	1.89%	3	3	0.9796	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.23.0 9.04	40.7	0.14126	21-Sep-04	0.333	2.60%	2	3	0.9718	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.14.1 0.04	42.9	0.14860	12-Oct-04	0.320	5.42%	3	2	0.9901	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	
FAL.3T3.SLS.04.11 .04	39.9	0.13836	2-Nov-04	0.259	2.51%	4	3	0.9816	100, 82.6, 68.3, 56.4, 46.7, 38.6, 31.9, 26.3	1.21	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
------------------------------	--------------------------------------	-----------	---	---	--	--	--	------------------------------------	-----------------------------------	---------------------------------	----------------------------------	----------------------------------

IIVS**Phase Ia**

B1	NA	NA	24-Aug-02	0.306	17.18%	1	0	0.5129	100, 56.2, 31.6, 17.8, 10, 5.63, 3.17, 1.78	1.78	NO	VC difference > 15%. No points between 50 & 90% viability.
B2	53.7	0.186	24-Aug-02	0.280	38.89%	1	0	0.3966	100, 56.2, 31.6, 17.8, 10, 5.63, 3.17, 1.78	1.78	NO	VC difference > 15%. No points between 50 & 90% viability.
B3	34.7	0.120	25-Aug-02	0.452	1.92%	0	1	0.9877	100, 56.2, 31.6, 17.8, 10, 5.63, 3.17, 1.78	1.78	NO	No points between 10 & 50% viability.
B4	34.2	0.119	25-Aug-02	0.428	4.07%	0	3	0.9664	100, 56.2, 31.6, 17.8, 10, 5.63, 3.17, 1.78	1.78	NO	No points between 10 & 50% viability.
B5	35.9	0.125	26-Aug-02	0.409	3.71%	0	1	0.9872	100, 56.2, 31.6, 17.8, 10, 5.63, 3.17, 1.78	1.78	NO	No points between 10 & 50% viability.
B6	39.0	0.135	26-Aug-02	0.382	0.09%	0	0	0.9649	100, 56.2, 31.6, 17.8, 10, 5.63, 3.17, 1.78	1.78	NO	No points between 10 & 90% viability.
B7	35.7	0.124	27-Aug-02	0.302	2.98%	0	2	0.9773	100, 56.2, 31.6, 17.8, 10, 5.63, 3.17, 1.78	1.78	NO	No points between 10 & 50% viability.
B8	36.1	0.125	27-Aug-02	0.299	6.86%	0	1	0.9792	100, 56.2, 31.6, 17.8, 10, 5.63, 3.17, 1.78	1.78	NO	No points between 10 & 50% viability.
B9	41.5	0.144	29-Aug-02	0.342	6.02%	1	1	0.9831	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B10	45.1	0.156	29-Aug-02	0.358	1.51%	1	1	0.9664	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B11	43.8	0.152	30-Aug-02	0.366	4.26%	1	0	0.9936	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	NO	No points between 50 & 90% viability.
B12	44.6	0.155	30-Aug-02	0.359	0.95%	1	1	0.9864	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B13	44.5	0.154	4-Sep-02	0.538	0.37%	1	1	0.9799	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B14	43.9	0.152	4-Sep-02	0.491	6.43%	1	1	0.9869	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B15	37.8	0.131	5-Sep-02	0.357	9.90%	1	1	0.9906	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B16	40.4	0.140	5-Sep-02	0.336	10.55%	1	1	0.9832	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B17	39.7	0.138	6-Sep-02	0.464	2.31%	1	2	0.9780	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B18	38.1	0.132	6-Sep-02	0.426	11.25%	1	1	0.9910	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B19	36.7	0.127	7-Sep-02	0.378	4.90%	1	1	0.9928	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
B20	36.5	0.127	7-Sep-02	0.354	12.49%	1	1	0.9954	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
B21	46.7	0.162	8-Sep-02	0.453	0.44%	0	2	0.9800	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	NO	No points between 10 & 50% viability.
B22	41.8	0.145	8-Sep-02	0.439	0.63%	1	1	0.9802	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10, 6.8	1.47	YES	
Phase Ib												
A1 Preliminary	41.1	0.143	15-Jan-03	0.389	8.42%	1	1	0.9890	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10.0, 6.81	1.47	YES	
B1	43.5	0.151	22-Jan-03	0.569	6.41%	1	1	0.9822	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10.0, 6.81	1.47	YES	
B2	44.8	0.155	29-Jan-03	0.514	2.88%	1	1	0.9830	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10.0, 6.81	1.47	YES	
B3	38.5	0.133	5-Feb-03	0.519	1.00%	1	1	0.9854	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10.0, 6.81	1.47	YES	
B4	49.4	0.171	12-Feb-03	0.548	10.23%	0	2	0.9770	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10.0, 6.81	1.47	NO	No points between 10 and 50%; IC50 out of range
B5	41.9	0.145	26-Feb-03	0.507	5.41%	1	1	0.9747	100, 68.1, 46.4, 31.6, 21.5, 14.7, 10.0, 6.81	1.47	YES	
Phase II												
A1	41.3	0.143	23-Jul-03	0.546	3.97%	1	3	0.9902	100, 66.7, 44.4, 29.6, 19.8, 13.2, 8.78, 5.85	1.50	YES	
B1	39.6	0.137	28-Jul-03	0.375	1.11%	1	5	0.9559	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.5	1.40	YES	
B2	38.8	0.135	29-Jul-03	0.529	5.36%	2	5	0.9711	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.5	1.40	YES	
B3	30.0	0.104	30-Jul-03	0.527	1.74%	1	4	0.9854	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.5	1.40	NO	IC50 out of range
B4	42.6	0.148	13-Aug-03	0.483	7.35%	1	5	0.9891	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.5	1.40	YES	
B5	39.1	0.136	16-Sep-03	0.510	6.44%	3	5	0.9568	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.5	1.40	YES	
B6	38.2	0.132	23-Sep-03	0.433	2.75%	1	5	0.9668	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.5	1.40	YES	
B7	38.9	0.135	24-Sep-03	0.479	2.49%	1	5	0.9710	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.5	1.40	YES	
B8	45.2	0.157	1-Oct-03	0.547	3.52%	1	5	0.9798	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.5	1.40	YES	
Phase III												
A1	42.1	0.146	3-Feb-04	0.429	3.86%	2	5	0.9691	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49,	1.40	YES	
A2	42.4	0.147	10-Feb-04	0.494	0.10%	2	4	0.9874	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49,	1.40	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
A3	41.0	0.142	17-Feb-04	0.458	1.06%	1	4	0.9858	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
A4	37.2	0.129	9-Mar-04	0.417	7.26%	1	4	0.9893	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
A5	33.0	0.114	23-Mar-04	0.346	1.01%	2	3	0.9758	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B1	45.9	0.159	26-Jul-04	0.399	0.81%	1	5	0.9709	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49,	1.40	YES	
B2	44.5	0.154	27-Jul-04	0.379	5.70%	3	4	0.9828	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49,	1.40	YES	
B3	40.1	0.139	28-Jul-04	0.344	14.50%	2	5	0.9364	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49,	1.40	YES	
B4	42.2	0.146	23-Aug-04	0.493	3.37%	1	3	0.9874	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49,	1.40	YES	
B5	47.2	0.164	24-Aug-04	0.485	7.64%	2	2	0.9864	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B6	46.1	0.160	28-Sep-04	0.462	1.12%	1	4	0.9824	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B7	40.7	0.141	1-Oct-04	0.372	10.21%	1	5	0.9808	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B8	41.2	0.143	4-Oct-04	0.427	0.90%	1	4	0.9826	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B9	43.4	0.150	12-Oct-04	0.413	4.72%	1	5	0.9758	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B10	43.7	0.151	13-Oct-04	0.465	2.54%	2	5	0.9833	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49,	1.40	YES	
B11	42.3	0.147	2-Nov-04	0.398	4.84%	1	3	0.9920	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B12	32.5	0.113	9-Nov-04	0.355	1.15%	1	3	0.9888	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B13	41.6	0.144	10-Nov-04	0.362	5.53%	1	4	0.9831	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B14	21.4	0.074	16-Nov-04	0.445	4.98%	3	3	0.9568	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	NO	IC50 out of range
B15	43.5	0.151	8-Dec-04	0.442	2.26%	1	3	0.9932	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B16	37.2	0.129	14-Dec-04	0.436	5.18%	1	5	0.9757	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B17	43.2	0.150	15-Dec-04	0.373	3.10%	1	3	0.9869	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49	1.40	YES	
B18	41.0	0.142	19-Jan-05	0.385	1.43%	1	3	0.9739	100, 71.4, 51.0, 36.4, 26.0, 18.6, 13.3, 9.49,	1.40	YES	

3T3 NRU SLS Data

Experiment I.D. ¹	IC50 (ug/mL) Decimal Format	IC50 (mM)	Test Chemical Application Date	Mean Vehicle Control (VC) OD ²	Difference of right/left VC from mean VC ³	Number of Points 0 - 50 % ⁴	Number of Points 50 - 100 % ⁵	R2 (from PRISM) ⁶	Dilutions (ug/mL) ⁷	Dilution Factor ⁸	Acceptable Tests ⁹	Rationale for Unacceptability
------------------------------	--------------------------------------	-----------	---	---	--	--	--	------------------------------------	-----------------------------------	---------------------------------	----------------------------------	----------------------------------

¹ Positive control test I.D.² Mean OD value for all VC wells in test plate³ Difference of right and left VC column of wells in the test plate⁴ % viability ICx values between 0 and 50% viability; test acceptance criterion; Phases Ia and Ib = number of points between 10 - 50%⁵ % viability ICx values between 50 and 100% viability; test acceptance criterion; Phases Ia and Ib = number of points between 50 - 90%⁶ calculated value from the Prism® software⁷ Test chemical concentrations from stock applied to the cells⁸ Step-wise dilution factor⁹ Determination whether test meets or doesn't meet test acceptance criteria

Shaded boxes identify values that do not meet the specific test acceptance criteria

Abbreviations:

NR = Neutral Red

R&D = Research and Development

PC = Positive Control

C1 - C8 = concentration series

NA = not available

2X = Two times the concentration that will be applied to the cells

VC = Vehicle Control

R2 = Coefficient of Determination

OD = Optical Density

Phase	ECBC PC Acceptance Range (ug/mL)	FAL PC Acceptance Range (ug/mL)	IIVS PC Acceptance Range (ug/mL)
Ib (3T3)	28.8 – 47.7	25.2 – 59.5	34.5 – 47.3
II (3T3)	26.4 – 56.3	31.5 – 54.9	33.6 – 50.6
III (3T3)	30.8 – 51.6	27.2 – 64.7	31.8 – 49.3